



# California Medical Device Recall Information



## Recall Name

### Tandem Diabetes Care Recalls Insulin Cartridges Due to Cartridge Leak

Recall Date	Product Description	Recalling Firm	Recall Reason
Initial: 01/10/14  Expanded: 01/20/14	Specific lots of insulin cartridges used with <b>t:slim® Insulin Pump</b> .  4,746 boxes of cartridges (10 cartridges per box).	<b>Tandem Diabetes Care, Inc.</b> San Diego, CA	<i>Affected cartridges may be at risk for leaking.</i>  <i>Leaking cartridges could result in the delivery of too much or too little insulin.</i>
Recall Class	Product Identification	Distribution	Affected Dates
I	Follow the link below for affected lot numbers:  <a href="#">Affected Lot Numbers</a>	<b>CA</b> , nationwide	Shipped on or after  May 17, 2013.

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm385837.htm>